CurAlert System Tester



PART NUMBER: **2519**

FEATURES

Tests Infant Transmitter operation and battery condition
Tests CurAlert Door Monitor operation
Displays CurAlert system alarms
Resets Door Monitor alarms
Helps Locate misplaced Infant Transmitters
Helps identify presence of RF interference
Low cost
Easy to use

TECHNICAL SPECIFICATIONS

9Vdc battery or 12Vdc plug pack operation
Operating frequency 433mHz
Two line LCD display
Red LED and audible alarm indication
Green LED test mode indication
Dimensions 130 x 70 x 25 mm
Durable ABS plastic enclosure with plastic membrane front panel

ACCESSORIES

2404 Plug pack, **2515** Infant Transmitter, **2573** Wristband Duress Transmitter, **2521** Door Monitor, **2558** Wireless Call Button, **2559** Wireless Reset Button

cura1.com

cura¹ products are designed to augment well designed care programs and practices, cura1 does not claim that any of its products are suitable as a substitute, or as a total solution for individuals that require care and constant supervision, and are at risk.