Central Monitor



PART NUMBER: **2555**

FEATURES

Wireless alarm monitor for use with other cura[1] wireless products
Indicates other cura[1] wireless equipment low battery and fault conditions
Easy installation and simple operation
Low maintenance and long battery life
Fully wireless - no cabling required
Automatic "read-in" programming
High visibility flashing LED light and adjustable audible alarm
Can be connected to a nurse call system
Caregiver key reset option

TECHNICAL SPECIFICATIONS

9-13.8 Vdc operation (stand by current consumption <100mA)
4 of D type Alkaline internal backup batteries
Battery life in AC mains fail 10 hours, in stand-by 2 to 3 years
Two line LCD backlit display
Red LED and audible sounder low battery indication
Operating frequency 433MHz
Volt free normally open alarm output
User selectable reset functions
Automatic "out of range" and "interference detection" with alert
Automatic "read-in" programming procedure
Uniquely identifies up to 200 - 2615 or 2608 Wireless Bed Monitors (FallAlert System)
Uniquely identifies up to 1000 - 2572 Resident Transmitters (DoorAlert System)
Uniquely identifies up to 16 - Door Locations (DoorAlert System)
Uniquely identifies up to 1000 - 2558 Wireless Call Buttons (StaffAlert System)
Uniquely identifies up to 1000 - 2573 Wristband Duress Transmitters (StaffAlert System)
Up to 8 unique area IDs
Dimensions 210 x 150 x 45 mm
White durable ABS plastic enclosure with plastic membrane front panel
Low battery and AC power loss detection and alert

ACCESSORIES

2615 Wireless Bed Monitor, **2608** Wireless Bed Monitor, **2404** Plug pack, **2401** 1A Power Supply, **2402** 2.5A Power Supply, **2403** 5A Power Supply, **2561** Wireless Indicator Light, **2562** Wireless Bumper, **2552** Door Monitor, **2572** Wrist Band Transmitter, **2573** Wrist Band Transmitter, **2558** Wireless Call Button, **2559** Wireless Reset Button, 2642 Caregiver Reset Key

cura1.com

PM003

cura[1] products are designed to augment well designed care programs and practices, cura1 does not claim that any of its products are suitable as a substitute, or as a total solution for individuals that require care and constant supervision, and are at risk.